

**2102 Rock Salt Moisture Balance Rapid Moisture Test****2102.1 Sampling – See 2101.1****2102.2 Moisture Balance Procedure**

The moisture balance used should have a capacity of 35g and a minimum readability of 0.005g. Set up the moisture balance according to the manufacturer's recommendations. Operate instrument in an automatic mode for the stop point and at 160 C drying temperature. Place approximately 25 g on the sample pan and press start. Record the % moisture at completion of the test.